



BREATHING MACHINE SIMULATED WEARING TREATMENT TESTER

Breathing Machine Simulated Wearing Treatment Tester ToronFM™ BMSWTT simulates the wearing pretreatment process before conducting respiratory resistance or filtration material penetrability tests on breathing machines.

The breathing machine is activated, the saturator turned on, and the system allowed to stabilize. The particle-filtering half mask being tested is mounted on a dummy head. Over the test period, the mask is removed and reattached every 20 minutes, totaling ten fittings during the test.

APPLICATIONS

Standard

- 1. EN 149 Section 8.3.1 – Specifies the total inward leakage (TIL) test method for filtering half masks to measure how much airborne particles leak into the mask during simulated human movement and breathing.


FEATURES

- 1. This machine includes a breathing simulator, dummy head, saturator, temperature controller, flowmeter, and more. It allows for adjustment to 25 cycles per minute at 2.0 liters per stroke. The air can be saturated at a temperature of **(37 ± 2)°C** at the dummy head's mouth.
- 2. **User-Friendly Operation:** The touch-screen interface ensures easy and efficient operation.

TECHNICAL SPECIFICATIONS

Model	ToronFMBM™ SWTT
-------	-----------------

Breathe Frequency	25 cpm
Tidal Volume	0.5L/min-2.8L/min
Flowmeter Range	5-45L/min, accuracy: 1L/min
Weight	50 kg
Voltage	AC220V 50Hz (110V is also available)



Torontech Inc.

251 Consumers Road, Suite 1200
Toronto, Ontario, M2J 4R3, Canada 

Phone: +1 416 368 2721 | Fax: +1 416 981 7652
Email: info@torontech.com

Torontech, LLC

601 Brickell Key Drive, Suite 700
Miami, FL 33131, USA 

Toll-Free: 1-866-383-7919
Email: sales@torontech.com

Torontech ME FZE

Meydan Grandstand, 6th Floor,
Meydan Road, Nad Al Sheba,
Dubai-261440, UAE 

Phone: +971 488 19252
Email: me@torontech.com

